

VA, Inc. dba Vincent Associates

Declaration of Conformity

Directive(s) EMC, LOW VOLTAGE, UL, CSA

Type of Equipment: Shutter Driver Equipment

Applied Standards:
(Vincent Associates has successfully completed and is compliant with the following tests performed at MET Laboratories, Inc.)

Application of Council Directive(s): EMC 89/336/EEC as amended by 93/68/EEC
Standards to which Conformity is Declared: 61326:1997 + A1:1998 + A2:2002 + A3:2003 Class A/B
UL61010A-1 First Edition; Electrical Equipment for Laboratory Use; Part 1: General Requirements, Rev. December 6, 2002
CSA C22.2 N0 1010-1 First Edition; Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use; Part 1: General Requirements, Rev. 1999
EN 61326-1: 2006 – Electrical Equipment for Measurement, Control, and Laboratory Use. EMC Requirements as follows:
EN 61000-4-2:1995 + A1:1998 + A2:2002 – Electrostatic Discharge Immunity
EN 61000-4-3:2006 – Radiated Electromagnetic Field Immunity
EN 61000-4-4:2004 – Electrical Fast Transient/Burst Immunity
EN 61000-4-5:2005 – Surge Immunity
EN 61000-4-6:2003 + A1:2003 + A2:2005 – Conducted Radio-Frequency Immunity
EN 61000-4-11:1994 – Voltage Dips and Interruptions
CISPR 16-2-3:2003 – Radiated Emissions – Class A
CISPR 16-2-1:2003 – Conducted Emissions – Class A

Model Number(s): VCM-D1

Manufacturer's Name and Address: Vincent Associates
803 Linden Avenue
Rochester, NY 14625

Based on the Declaration of Conformity Certificates issued by our test laboratory, MET Laboratories, Inc., I declare that the equipment specified above conforms to the listed directives and standards.

Place: Rochester, NY

Original Date: April 28, 2004

Updated: January 9, 2009



Signature

Stephen T. Pasquarella
Full Name

Sr. Vice President
Title

