

VA, Inc. dba Vincent Associates®

Declaration of Conformity

Directive(s) EMC, LOW VOLTAGE, UL, CSA

Type of Equipment: Shutter Driver Equipment

Applied Standards:
(Vincent Associates has successfully completed and is compliant with the following tests performed at MET Laboratories, Inc.)

Application of Council Directive(s): EMC 89/336/EEC as amended by 92/31/EEC and 93/68/EEC

Standards to which Conformity is Declared: 61326:1997 + A1:1998 + A2:2001 + A3:2003 Class A

UL 61010-1/CSA C22.2 No. 61010-1, Second Edition: Safety of Electrical Equipment for Measurement, Control and Laboratory Use, Rev. October 28, 2008.

Title 47 of the Code of Federal Regulations (CFR), Part 15 Subpart B for a Class A digital device.

15.107 (a) Conducted Emission Limits for a Class A Digital Device

15.109 (a) Radiated Emission Limits for a Class A Digital Device

EN 61326-1: 2006, Electrical Equipment for Measurement, Control, and Laboratory Use.

EN 61000-4-2:2001 – Electrostatic Discharge Immunity

EN 61000-4-3:2006 – Radiated Electromagnetic Field Immunity

EN 61000-4-4:2004 – Electrical Fast Transient/Burst Immunity

EN 61000-4-5:2006 – Surge Immunity

EN 61000-4-6:2007 – Conducted Radio-Frequency Immunity

EN 61000-4-11:2004 – Voltage Dips, Interruptions and Variations

CISPR 11:2007 – Radiated Emissions – Class A

CISPR 11:2007 – Conducted Emissions – Class A

Model Number(s): VDM1000

Manufacturer's Name and Address: Vincent Associates
803 Linden Avenue
Rochester, NY 14625

Based on the Declaration of Conformity Certificates issued by our test laboratory, MET Laboratories, Inc., I declare that the equipment specified above conforms to the listed directives and standards.

Place: Rochester, NY

Original Date: September 28, 2009

Updated:



Signature

Stephen T. Pasquarella
Full Name

Sr. Vice President
Title

