Uniblitz® LS3

3mm Uni-Stable Laser Shutter

Overview

The Uniblitz LS3 is a high-performance laser shutter, ideal for precision exposure control, laser switching, and other applications that benefit from the precise, repeatable characteristics of the LS shutter series. These include low level laser chopping, and pulse gating, selection, and modulation to 200Hz. Uni-stable shutter devices, like the LS3, require power to hold the blades in the open state (though the device can be custom-configured for "normally open" operation).

Specifications

Electrical Specifications	
Coil resistance	48 OHMS
Voltage to Open	+65 VDC
Hold Voltage (Nominal) ¹	+5 VDC

Mechanical Specifications

Weight Housed	210.0 g
Operating Temp.	0 - 80 °C
Max. Opening Bounce	15%
Max. Closing Bounce	5%
Max. Freq. of Operation ²	50 Hz / 200 Hz
Number of Shutter Blades	2

¹*Voltage level required across actuator coil when being held in the open position.*

² (Continuous/Burst) Continuous frequency rating specified at shutter's minimum exposure pulse. Burst frequency rating specified for four (4) seconds maximum with one (1) minute minimum between bursts.


Key Features

- 3mm aperture
- Laser Energy Rating: 5 W/mm² (with "ZM" blades)
- Configured for the <u>VCM-D1</u>
 <u>Shutter Driver</u>
- RoHS Compliant
- Transfer time on opening:
 0.5 milliseconds
- Transfer time on closing:0.6 milliseconds



Need Support? Please visit our website or email us at **info@uniblitz.com**. Tel: +1 585-385-5930 | Fax: 585-385-6004 | 803 Linden Ave. Rochester, NY 14625 Updated 5/25 | Document Version 6.1 | ©2025 Vincent Associates

Shutter Timing Data



¹ Under no circumstances should any type of lubricant be applied to the shutter blade area. Lubricating the shutter blades will likely slow the shutter down and may eventually render it inoperable.

/ VCM-D1 and "T" blades) ¹	Time (msec.)	
Delay time on opening after current ap	plied 1.0	
Transfer time on opening	0.5	
Total opening time	1.5	
Min. dwell time with min. input pulse	0.8	
Min. eqivalent exp. time	1.5	
Transfer time on closing	0.6	
Total window time	1.9	
Min. exposure time	2.0	
Typical exposure pulse	>2.3	
	Delay time on opening after current ap Transfer time on opening Total opening time Min. dwell time with min. input pulse Min. eqivalent exp. time Transfer time on closing Total window time Min. exposure time	Delay time on opening after current applied1.0Transfer time on opening0.5Total opening time1.5Min. dwell time with min. input pulse0.8Min. eqivalent exp. time1.5Transfer time on closing0.6Total window time1.9Min. exposure time2.0

Electronic Sync.

The electronic synchronization system provides a feedback signal (through the driver utilized) after the shutter transfers to the open state. The system incorporates an infrared emitting diode, an infrared sensitive detecting transistor, and an interrupting vane. The vane is attached to the shutter so as to block the light path between the emitter and detector in the closed position. When the shutter transfers to the 80% open position, the vane is removed from the infrared light path, allowing the emitter to switch the detector to the active state. For the LS3, this system uses a similar activation flag attached to the mechanism, which triggers a reflective emitter/ detector device. No connection to the designated synchronization pins when no electronic sync. is selected.

Product Options

LS3 23456-7-8


1 Shutter Series: **2** Driver Compatibility: **B** Housing: 4 Blades: ² LS3 S: Use with VCM-D1, VMM-D3/D4 T: Low Energy (Teflon[®]) 1: Un-Housed **ZM:** High Energy (AIMgF2)³ **E:** Use with V880, VED24, or D880C 2: #2 Housing **Z:** High Energy (AlSiO)³ **D** Encapsulated Coil: 8 Mount: ⁴ **6** Electronic Sync: **6** Connector: 0: Omitted EC: Included **21:** Zeiss Axiovert **105:** C-Mount (Male) L: 18" Flying Leads 1: Included Leave blank for 7-pin Leave blank if not 24: Olympus BX/IX 106: C-Mount (Female) Wire Pro connector required **27:** Nikon *Leave blank if not required*

² Other blade coating options may be available by special order.

³ Input side only; Teflon[®] coating is on opposite side to protect shutter blade surface. Light source must be input to the reflective side only.

⁴ Mounts are only compatible if #2 housing is optioned as well.


Ex: LS3S2ZM0-EC-21




Uniblitz® LS3 Technical Drawings

Uniblitz® LS3 Technical Drawings


Un-Housed



501A-S7 Adapter Cable¹ (Not included)



Connector Layout





¹ 501A-S7 Shutter Adapter Cable is required for use with VMM-D3, VMM-D4, D880C and any VCM-D1 purchased prior to November 2022.